



Guru Gobind Singh Indraprastha University
Sector 16-C, Dwarka, New Delhi-110 078
(Purchase Branch)

Ph: 011-25302149-150 Email: purchasebranch@ipu.ac.in
Website: www.ipu.ac.in



F.No.15/GGSIPU/CEPS/2023/ 52

Dated. 30.04.2024

NOTICE INVITING QUOTATION

Subject: Inviting Quotations for procurement of 01 nos. of Laboratory Refrigerator i.e. -20° Refrigerator for CEPS, GGSIP University.

The University is in process of procurement of **Laboratory Refrigerator i.e. -20° Refrigerator**, as per details below:-


S. No.	Item Description/Specifications	Qty.	Rate per Unit (in Rs.)	Total Amount (in Rs.)
i.	<ul style="list-style-type: none">Refrigerator must have (1) hinged solid doors.Refrigerator must have positive forced-air circulation.Refrigerator must have automatic defrost system.Refrigerator must have Microprocessor based Electronic Controller with 7 segment LED display.Refrigerator must have mercury-free LED interior lighting.Refrigerator must have externally mounted internal light switch.Refrigerator must have a capacity of 332 Liters capacity.Refrigerator must have a 1-minute door opening recovery of less than or equal to 8 minutes.Refrigerator must have set of four casters standard and installed in factory.Refrigerator must have digital temperature display.Refrigerator must have temperature resolution to 0.1°C.Refrigerator must have a temperature uniformity of 1.3°C or better.Refrigerator must have adjustable cold and warm alarms.Refrigerator must have both audible and visual alarms.Refrigerator must have door ajar alarm and icon.	01 Nos.		

S. No.	Item Description/Specifications	Qty.	Rate per Unit (in Rs.)	Total Amount (in Rs.)
	<ul style="list-style-type: none"> • Refrigerator must have service alarm and icon. • Refrigerator must have low battery alarm. • Refrigerator must have (4) internal wire shelves. • Refrigerator must have standard 1" rear access port standard. • Refrigerator must have standard remote alarm terminal standard. • Refrigerator must be CE certified – Refrigerator must incorporate Rodent proof power cord. • Refrigerator must use environmentally friendly SNAP & F-GAS Compliant HC refrigerants. • R Refrigerator must operate at less than or equal to 53.5 dba. • Refrigerator must support a standard settable range of -25°C to -15°C. • Refrigerator must include soft touch grip for full door length handle. 			
Total Amount (in Rs.)				
GST (as applicable)				
Grand Total (in Rs.)				

The interested bidders are requested to submit the quotation (in sealed envelopes) on the above mentioned format, in the office of undersigned latest by **6th May, 2024 at 12:00 Noon.**


Terms & Condition

- i. FOR : Guru Gobind Singh Indraprastha University, Sector-16/C, Dwarka, New Delhi.
- ii. Payment : After satisfactory receipt of the material as per the specification
- iii. Delivery : Within 5 days after issuance of Purchase Order.
- iv. Material will be delivered along with triplicate copies of Challans, Bills and Purchase Order to "Assistant Registrar, University Central Store, Ground Floor, "A" Wing, Administrative Block, GGSIPU, Sector -16-C, Dwarka, New Delhi (Ph: 011-25302141,44)" between 10.00 AM to 04.00 PM.



(Shikha Agarwal)
Deputy Registrar (Purchase)

To,
In-Charge, UITS with a request to upload on University Website.



(Rajesh Sharma)
Assistant Registrar (Purchase)